

140ACT GENERAL NOTES:

1. TOTAL ESTIMATED DESIGN WEIGHT: () LBS.
2. PANEL CONSTRUCTION: NOISELOCK V (INNER WALLS), NOISELOCK IV (OUTER WALLS) AND NOISELOCK VI (INNER & OUTER CEILING)

INTERIOR SURFACES: 22GA. GALVANIZED STUCCO EMBOSSED PERFORATED STEEL
EXTERIOR SURFACES: 14GA. COLD ROLLED STUCCO EMBOSSED SOLID STEEL
FLOOR INTERIOR SURFACE: 11GA. HOT ROLLED STEEL
3. FINISH: BAKED ACRYLIC ENAMEL PAINT
COLOR: CHOICE OF (10) STANDARDS
4. FLOOR COVERING: INDUSTRIAL NYLON CARPET
COLOR: (TO MATCH PAINT COLOR SELECTION)
3/8" THICK UNDERLAYMENT AND 2 1/2" H BLACK COVE BASE MOLDING INCLUDED.
5. DOOR: (3) 2' X 9" WIDE X 6' – 1 1/2" HIGH CLEAR OPENING CAM-LIFT SPLIT-FRAME WITH PUSH PLATE AND PULL HANDLES.
NOTE: DIMENSIONS LOCATING DOOR ARE SHOWN TO THE CLEAR OPENING
6. WINDOW: (2) 2'6" WIDE X 2'10" HIGH ALUMINUM FRAME DOUBLE GLAZED WITH 1/4" THICK LAMINATED SAFETY GLASS.
7. VIBRATION ISOLATOR RAILS TO BE PROVIDED BELOW FLOOR PANELS.
8. 2" MINIMUM CLEARANCE IS REQUIRED BETWEEN THE EXTERIOR OF TESTING SUITE AND HOST FACILITY WALLS/CEILING.
9. DUST SEAL TO BE PROVIDED WITHIN AIR SPACE AROUND DOORS AND WINDOWS. SILL ANGLE TO BE PROVIDED BETWEEN DOORS.
10. LARGEST PANEL SIZE: () WIDE X () LONG X () THICK. CUSTOMER TO ENSURE ADEQUATE SITE ACCESS.
11. VENTILATION: () CFM SYSTEM (EXAM) AND () CFM SYSTEM (CONTROL) COMPRISED OF INTEGRATED SPASAVER EXHAUST/INLET AIR CEILING SILENCER PANELS AND 120V, 60HZ, 14W FANS.

INDUSTRIAL ACOUSTICS COMPANY, INC.

Rev. 1 May 2006

12. ELECTRICAL:
- () 120V RECESSED INCANDESCENT LIGHT FIXTURE
 - (4) 120V TOGGLE SWITCH FOR FANS AND LIGHTS
 - () 120V DUPLEX RECEPTACLE
 - (2) 120V QUAD RECEPTACLE WITH POWER CORD (ROOF MOUNTED)
 - () 120V J-BOX WITH POWER CORD (ROOF MOUNTED)
 - (1) JP-81 AUDIOMETRIC JACK PANEL COMPRISED OF:
 - (10) SWITHCRAFT #12B JACKS (INSIDE AND OUTSIDE)
 - (1) CINCH-JONES #S303-RP (INSIDE)/(1) CINCH JONES #P-303-RP (OUTSIDE)
 - (1) CINCH-JONES #S304-RP (INSIDE)/(1) CINCH JONES #P-304-RP (OUTSIDE)
 - (2) 1" I.D. GROMMETED OPENINGS (INSIDE AND OUTSIDE)
 - (LOT) BOXES, CONNECTORS, ETC. AS REQUIRED FOR COMPLETE INSTALLATION

NOTE:

- A. CUSTOMER TO PROVIDE 110V, 60 HZ, 20 AMP POWER TO ROOF MOUNTED QUAD RECEPTACLE POWER CORD.
 - B. ALL ELECTRICAL IS WIRED PER 2002 NATIONAL ELECTRIC CODE
 - C. BOOTH IS FACTORY UL CERTIFIED AND LABELED
13. CUSTOMER TO PROVIDE A FINISHED FLOOR WHICH IS WITHIN +/- 1/4" OF LEVEL IN 10 NON-CUMULATIVE FEET.