

**40M GENERAL NOTES:**

1. TOTAL ESTIMATED DESIGN WEIGHT: ( ) LBS.
2. PANEL CONSTRUCTION: NOISELOCK V (WALLS) AND NOISELOCK VI (CEILING)  
  
INTERIOR SURFACES: 22GA. GALVANIZED STUCCO EMBOSSED PERFORATED STEEL  
EXTERIOR SURFACES: 14GA. COLD ROLLED STUCCO EMBOSSED SOLID STEEL  
FLOOR INTERIOR SURFACE: 11GA. HOT ROLLED STEEL
3. FINISH: BAKED ACRYLIC ENAMEL PAINT  
COLOR: CHOICE OF (10) STANDARDS
4. FLOOR COVERING: INDUSTRIAL NYLON CARPET  
COLOR: (TO MATCH PAINT COLOR SELECTION)  
3/8" THICK UNDERLAYMENT AND 2 1/2" H BLACK COVE BASE MOLDING INCLUDED.
5. DOOR: (1) 2' X 9" WIDE X 6' – 1 1/2" HIGH CLEAR OPENING CAM-LIFT SPLIT-FRAME WITH 1' 4" W X 2' 0" H ALUMINUM FRAME DOUBLE GLAZED WINDOW, PUSH PLATE AND PULL HANDLES. NOTE: DIMENSIONS LOCATING DOOR ARE SHOWN TO THE CLEAR OPENING
6. VIBRATION ISOLATOR RAILS TO BE PROVIDED BELOW FLOOR PANELS.
7. 2" MINIMUM CLEARANCE IS REQUIRED BETWEEN THE EXTERIOR OF TESTING BOOTH AND HOST FACILITY WALLS/CEILING.
8. DIVIDER CURTAINS AND STOOLS PROVIDED FOR EACH STATION.
9. LARGEST PANEL SIZE: ( ) WIDE X ( ) LONG X ( ) THICK. CUSTOMER TO ENSURE ADEQUATE SITE ACCESS.
10. VENTILATION: ( ) CFM SYSTEM COMPRISED OF INTEGRATED SPASAVER EXHAUST/INLET AIR CEILING SILENCER PANELS AND 120V, 60HZ, 14W FANS.
11. ELECTRICAL:  
( ) 120V RECESSED INCANDESCENT LIGHT FIXTURE  
(2) 120V TOGGLE SWITCH FOR FANS AND LIGHTS  
(1) 120V QUAD RECEPTACLE WITH POWER CORD (ROOF MOUNTED)  
(1) AUDIOMETRIC JACK NETWORK COMPRISED OF:

**INDUSTRIAL ACOUSTICS COMPANY, INC.**

Rev. 1 May 2006

- (1) MASTER JACK PLATE WITH ( ) SWITCHCRAFT #112B 3-  
CONDUCTOR JACKS (EXTERIOR)
- ( ) STATION JACK PLATES WITH (3) SWITCHCRAFT #112B 3-  
CONDUCTOR JACKS (INTERIOR)
- (LOT) BOXES, CONNECTORS, CABLE, WIREMOLD, ETC. AS REQUIRED  
FOR COMPLETE INSTALLATION

**NOTE:**

- A. CUSTOMER TO PROVIDE 110V, 60 HZ, 20 AMP POWER TO  
ROOF MOUNTED QUAD RECEPTACLE POWER CORD.
  - B. ALL ELECTRICAL IS WIRED PER 2002 NATIONAL ELECTRIC  
CODE
  - C. BOOTH IS FACTORY UL CERTIFIED AND LABELED
12. CUSTOMER TO PROVIDE A FINISHED FLOOR WHICH IS WITHIN +/- ¼" OF  
LEVEL IN 10 NON-CUMULATIVE FEET.